

Global Declaration of product release Certification of conformity

Manufacturer:

Equflow B.V.
Voorschakelstraat 8
5349 CC Oss
The Netherlands

Declares that:

Description	PVDF flow tube 8.5 for Tubeholder / Clamp - 12 mm hose barb
Part number	0085PH000TX-CK001
Production Lot number	579-100518
Serial numbers	0156766-0156839
Lot release date	22-Apr-2025
Certificate date	04-Jun-2025
Expire date (*)	22-Apr-2028

Is / are manufactured, tested and calibrated individually in compliance with the manufacturer production and quality directives. The products were submitted by the manufacturer to the factory production control department that has performed the outgoing inspection.

Certificate number: GC00579-0085PH000TX-CK001
Accreditation by Equflow QA department



Date: 04-Jun-2025
Released by: Niek Smeenk
Function: QA & Product manager

Note:

To receive a copy of applicable Quality directives or part of it, description of the calibration or test procedure, or any other document in relation to validation or certification, contact your provider or send a request to sales_sensors@saint-gobain.com to the attention of the QA manager.

(*) Expire date is 36 months after the lot release date, if the flowmeters are stored properly**.

When sterilization with gamma radiation is used, the expire date will be 24 months after gamma irradiation, if sterilization is conducted within 12 months of the lot release date. If the sterilization is conducted more than 12 months after lot release, the expire date will be 36 months after lot release. In both cases the maximum gamma dose of 50 kGy must not be exceeded.

(**) Keep the materials sealed during storage, and avoid extreme temperatures and direct sunlight.